

REQUIREMENTS FOR SITE INSTALLATION


OF THE



COMPARATOR  
COMPUTER

STATINTL

A. Equipment Site

1. The equipment site must be 40 feet in length and a minimum of 18 feet in width as presented in a modified layout drawing; 

STATINTL

2. An additional space of 18 feet in length and 120 feet in width will be required for the machine room. This room will house an air handler, electric power, vacuum pumps, dry air supply, and air compressors. This space must be adjacent to the stereo comparator in order to have the shortest distance from the instrument for vacuum and air lines. This short distance is required to reduce friction and condensation, and by honoring a three second time element required for pressure in the vacuum and compressed air systems.

3. The site should have a minimum of 12 foot ceiling height in order to accomodate a 3 foot raised floor, permit an 8 foot finished ceiling height, and a minimum of 1 foot above the dropped ceiling for utilities.

4. Equipment must be used in a vaulted area

B. Equipment Size and Weight

1. The total instrument consists of seven major components as follows:

- a. The engine
- b. Utility Cabinet
- c. Card Punch
- d. Teletypewriter
- e. Electric Cabinet No. 1
- f. Electric Cabinet No. 2
- g. Electric Cabinet No. 3

Declass Review by NIMA/DOD

2. The engine is 14 feet in length, 7 feet in width, 5 3/4 feet in height.  
Total weight - 28,000 lbs.

a. The utility cabinet is 3 feet 10 inches in length, 2 feet 8 inches in width, 7 feet in height.  
Total weight - 1000 lbs.

b. The card punch is 2 feet 7 inches in length, 2 feet 4 inches in width, 2 feet 6 inches in height.  
Total weight - 290 lbs.

c. Teletypewriter is 2 feet 8 inches in length, 2 feet 4 inches in width, 2 feet 6 inches in height.  
Total weight - 125 lbs.

d. Electric Cabinet No. 1 is 3 feet 10 inches in length, 2 feet 8 inches in width, 7 feet in height.  
Total weight - 1000 lbs.

e. Electric Cabinet No. 2 is 3 feet 10 inches in length, 2 feet 8 inches in width, 7 feet in height.  
Total weight - 1000 lbs.

f. Electric Cabinet No. 3 is 3 feet 10 inches in length, 2 feet 8 inches in width, 7 feet in height.  
Total weight - 1000 lbs.

3. Total weight of instrument installed - 32,415 lbs.

#### C. Air Conditioning Requirements

1. The air conditioning system for this instrument is very complex. For identification purposes we will identify the air conditioning requirements as room air and equipment air.

##### a. Room Air

Room air will consist of two separate systems. A horizontal lammer flow at the position of the engine and a vertical lammer flow system in and around the remainder of the equipment.

##### b. The Instrument Air

Direct connected air will be required to the utility cabinet, electric cabinets Nos. 1, 2, and 3. In addition the engine will require four separate and direct equipment tie ins for lamp cooling, etc.

##### c. The Average Temperature

The average temperature requirement is 72°F, plus or

minus 0.5 and the relative humidity is 55 percent, plus 15 percent, minus 5 percent.

d. The Air CFM Requirement

1. Engine	1300 CFM
2. Utility Cabinet	450 CFM
3. Card Punch	220 CFM
4. Teletypewriter	220 CFM
5. Electric Cabinet No. 1	2000 CFM
6. Electric Cabinet No. 2	1300 CFM
7. Electric Cabinet No. 3	2400 CFM

e. Total CFM for complete instrument 7890 CFM

D. Electric Power Requirements

a.	1. Engine	110 VAC	50 amps maximum	30 amps cont.
	2. Utility Cabinet	110 VAC	10 amps	
	3. Card Punch	110 VAC	5 amps	
	4. Teletypewriter	110 VAC	5 amps	
	5. Electric Cabinet No. 1	110 VAC	90 amps maximum	50 amps cont.
	6. Electric Cabinet No. 2	110 VAC	40 amps maximum	30 amps cont.
	7. Electric Cabinet No. 3	110 VAC	55 amps	

b. Total electric requirement:

110 VAC 245 maximum amps 120 amps cont.  
Circuit shall be 4 wire, 3 phase, positive ground.

c. Thirty-two pair wire and conduit to the main computer room.

E. Vacuum

a. 24 CFM @ 0"Hg with a maximum three second delay pressure recovery.

F. Compressed Air

Firm requirements not available.